



Product Name : Medical Air Compressor

Product Model No : nice 5001

Trouble Shooting

**Doc no:
73-00-046**

1. General System Failure

SI. No.	Problem	Cause	Remedy
1	Power failure LED is ON with Audio and Visual Indication	Power Failure	Check the unit is Plugged in to Main supply
			Check the mains are switched ON.
			Otherwise Contact nice Neotech
2	System Failure LED is ON with Audio and Visual Indication	System Failure	Contact nice Neotech.